

AXIOM Gesellschaft für Diagnostica und Biochemica mbH	COVID-19 Antigen Rapid Test	2101	<i>Retrospective in vitro study</i>	98% sensitivity 100% specificity NP/Nasal swab	DE <sup>[2]</sup>	Nucleo-protein	Nasal swab, Nasopharyngeal swab, Throat swab		10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (PEI): Sensitivity of 100% at Ct $\leq$ 25; Manufacturer specificity: 100%						